

To our customers

Ref : FESN-2501
End of April 2025
Fuji Electric Co.,Ltd.

Fuji Electric Marine Exhaust Gas Cleaning System (EGCS)

Regarding Routine Maintenance of the Laser Gas Analyzer (ZQS)

We are pleased to hear of your continued prosperity.

We sincerely thank you for your continued patronage of our Marine Exhaust Gas Cleaning System (EGCS).

We would like to reiterate the routine maintenance procedures :

○ Areas requiring regular cleaning

- Inside the sampling section : Male Run T
- Inside the external drain pot : Ejector

- ※ Please perform cleaning when the gas analyzer triggers an alarm (Flow low).
Cleaning cycle : Every 3 to 7 days (depending on scrubber operation conditions).

○ Parts that need replacement

- Inside the sampling section : Filter and Pre-filter

- ※ If errors (Flow low) frequently occur in the gas analyzer and
the above cleaning does not improve the situation, please replace the filter.

Instruction Manual	
Male Run T	Male Run T Cleaning Instructions(Extraction unit)
Ejector	Exhaust gas analyzer ejector cleaning procedure
Filter and Pre-filter	Exhaust gas analyzer filter replacement procedure

Contact for inquiries regarding this matter

Fuji Electric Co., Ltd. Industry Business Group

Field Service Division, West Japan Customer Engineering Center

Kansai Area CE Department II , CE Sec.Ⅲ

x-kansai-ce@fujielectric.com

Sincerely